

DEMKO CERTIFICATE

Certificate No.	D-06358-M2
Page	1/3
Date of Issue	2019-10-30
Certificate Holder	Volex (Asia) Pte Ltd 35 TAMPINES ST 92 SINGAPORE, 528880 Singapore
Manufacturer	Volex (Asia) Pte Ltd 35 TAMPINES ST 92 SINGAPORE, 528880 Singapore
Production site	Volex Cable Assembly (Shenzhen) Co Ltd VOLEX BLDG 1173 SHENHUI RD BAO'AN, HENGGANG SHENZHEN, 518115 GUANGDONG China
	See Page 2
Certified Product Model	Non-rewirable two-pole hospital plug with earthing pin DM133SH
Trademark	Volex
Rated Voltage / Frequency	250 Vac
Rated Current / Power	10, 13 or 16 A
Insulation Class	-
Degree of protection (IP)	20
Tested acc. to	DS 60884-2-D1:2017
Test Report No.	4789149927 issued on 2019-10-29
Additional	The report was revised to include technical modifications.
Expire date	2028-01-30

Certification Manager
Jan-Erik Storgaard

This is to certify that representative sample(s) of the Product described herein ("Certified Product") have been investigated and found in compliance with the Standard(s) indicated on this Certificate, in accordance with the D Mark Requirements. The Designated Certificate holder is entitled to use the d or for cables <DEMKO> for the Certified Product manufactured at the production site(s) identified above, in accordance with the D Mark Service Agreement, including without limitation the D Mark Testing and Certification Services Service Terms. Only those Products bearing the D Mark should be considered as being covered by UL's D Mark Service. This Certificate shall remain valid through the expiration date, unless terminated earlier in accordance with the Service Agreement including without limitation if the Standard identified on this Certificate is amended or withdrawn prior the expiration date.

Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com
www.ul-europe.com



Appendix DEMKO CERTIFICATE

Certificate No. D-06358-M2
Page 2/3
Date of Issue 2019-10-30

Production Sites:

Volex Cable Assembly (Zhongshan) Co Ltd
2 Xingda St
Torch Development Zone
ZHONGSHAN, 528437 GUANGDONG China

Additional Information:

This certificate replaces previous certificate no. D-06358-M1 issued on 2018-07-16, due to :

1. adding new cable size
2. adding rated current 16A
3. additional label

For Class I equipment

Standard Sheet: DK 2-8a

Type of cord:

H05VV-F 3G 1.0-1.5 mm²(two cable suppliers)

(H)05VV-F 3G 1.0-1.5 mm²(one cable supplier)


Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com
www.ul-europe.com



Appendix DEMKO CERTIFICATE

Certificate No. D-06358-M2
Page 3/3
Date of Issue 2019-10-30



Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com
www.ul-europe.com

